

Manufacturer:



Smart Denture Conversions, LLC
 56 Hunter St
 Suite 320
 Apex NC, 27511
 855-550-8787
www.SmartDentureConversions.com

Training:

The following descriptions are insufficient to allow immediate use of the Smart Denture Conversions’ system. Knowledge of implant-prosthetic treatment and instruction in the handling of the Smart Denture Conversions’ system provided by an operator with the relevant experience are necessary. It is strongly recommended that new and experienced users of Smart Denture Conversions’ products complete special training before using a new product for the first time. Smart Denture Conversions offers a variety of training options. Please go the www.SmartDentureConversions.com for more information.

Product Description:

Prosthetic screws are pre-manufactured dental screws designed for fixing prosthetic restorations onto the implant abutment. An assortment of Prosthetic Screws is available for use with different prostheses or implant system components, depending on the dental implant platform or connection type. The table below summarizes the items:

Name	Part Number	Compatible TiBase	Driver	Material	# of Uses
Prosthetic Screw, Unigrip	PS-001	SF-003/SF-004	Nobel (Unigrip)	Titanium alloy - Ti-6Al-4V (90% Ti, 6% Al, 4% V)	Single
Prosthetic Screw, Straumann	PS-002	SF-003/SF-004	Straumann (SDS)		
Prosthetic Screw, TiLobe (Quad Drive)	PS-003	SF-003/SF-004	TiLobe		
Prosthetic Screw, Paltop (IHex)	PS-004	PT-003/PT-004	0.05” Hex		
Prosthetic Screw, BioHorizons	PS-005	SF-003/SF-004	0.05” Hex		
Prosthetic Screw, BioHorizons	PS-006	SF-003/SF-004	0.05” Hex		
Prosthetic Screw, TSV	PS-007	SF-003/SF-004	0.05” Hex		
Prosthetic Screw, Low Profile	PS-008	SF-012/ SF-012L	0.048” Hex		

Indications for Use:

The supplied Prosthetic Screws are indicated for use with Smart Denture Conversions’ TiBases and screw-retained multiple-unit abutments in the maxilla and mandible. The prosthetic screws can be used with either the Standard TiBases or Tall TiBases within their respective family. The supplied Prosthetic Screws are intended for single use and a single patient. Re-use of single use devices creates a potential risk of patient or user infection and misfitting components. For more specific information on process steps, please refer to the Technique manual located on the website www.SmartDentureConversions.com.

Contraindications:

It is contraindicated to using Smart Denture Conversions’ Prosthetic Screws in:

- Patients who are medically unfit for an oral surgical procedure.

Prosthetic Screw

- Patients in whom adequate sizes, numbers or desirable positions of implants are not reachable to achieve safe support of functional or eventually parafunctional loads.
- Patients who show signs of allergy or hypersensitivity to the chemical components of the materials listed in the chart above.

Warning:

- For immediate load prostheses screw manually, avoiding excessive torque, and prevent the implant from turning while tightening.
- When transferring to the patient, do not use the same screw that was used in the laboratory.
- Make sure the correct model of screw is used for each case.
- Components are to be used by dental health care professionals and are to be used in patients subject to dental implant treatment.
- If the indication or the nature of use is not clear, do not use until all points have been clarified.
- Do not use if package is damaged.
- Always inspect components before use. Do not use damaged, deformed, corroded, or discolored components.
- Overtightening may cause the Prosthetic Screw or other components to become deformed, broken, stuck or dislodged in the implant analog or abutment, resulting in damage to components.
- Ensure products are secured against aspiration when handled intraorally. Aspiration of products may lead to infection or unplanned physical injury.
- Failure to follow the procedures outlined in these instructions may lead to any or all of the following complications: Aspiration or swallowing of a component, follow-up treatment, incorrect impression resulting in incompatible restorations.
- As the clinical outcome of dental treatment is influenced by multiple variables, even if the product is used according to the instructions for use the residual risks described below can. anaphylaxis (severe allergic reaction); aspiration or swallowing of components; pain; local infection; inflammation; local irritation; loss of product function; follow-up treatment.
- Smart Denture Conversions is not liable for damage resulting from use outside the intended use of the product.

Cautions/Precautions:

The following precautions are required before or during treatment:

- Do not use Smart Denture Conversion components after the expiration date on packaging (if applicable).
- All products intended for single use must not be reused. Re-use of single use devices creates a potential risk of patient or user infection and misfitting components.
- Before every procedure make sure that all required components, instruments, and auxiliary equipment are complete, in operating order and available in the required quantity.
- If, due to unfavorable anatomical conditions, instruments do not fit or cannot be used for other reasons, the course of treatment planned with them must not be continued and alternatives must be sought.
- Always wear suitable personal protective equipment for your own safety.
- Position the patient such that the danger of aspiration of components is minimized.
- All components used in the patient's mouth must be secured to prevent aspiration and swallowing.

Magnetic Resonance Imaging (MRI) Safety Information:

All Smart Denture Conversion LLC products which remain in the patient's body have not been evaluated for safety and compatibility in the MR environment. They have not been tested for heating, migration, or image artifact in the MR environment. The safety of the Smart Denture Conversion LLC products in the MR environment is unknown. Scanning a patient who has such a product may result in patient injury.

Prosthetic Screw

Cleaning and Sterilization Instructions:

Components are delivered non-sterile by Smart Denture Conversions and are intended for single use. Prior to use, the devices must be cleaned and sterilized by the user.

Sterilization is to be performed corresponding to the following scheme:

1. **Preparation for sterilization:** Place components in a sterilization pouch which is legally marketed (for the US market: FDA-cleared) for use with the recommended sterilization parameters. Every sterilization package must have a sterilization indicator and sterilization date.
2. **Sterilization:**

Method	Cycle	Temperature	Exposure Time*	Dry Time	Cooling Time	Reference
Steam	Dynamic Air Removal (Prevacuum)	132°C (270°F)	4min	20min	30min	ANSI/AAMI TIR12:2010
Steam	Gravity Displacement	121°C (250°F)	30min	30min	30min	
Steam	Gravity Displacement	132°C (270°F)	15min	30min	30min	

*Minimum exposure times, the operating times are longer and may vary depending on the device.

Storage, Handling and Transportation:

The devices must be stored in a dry place in their original packaging at room temperature and protected from direct sunlight. Improper storage may compromise essential material and design characteristics, leading to device failure.

Disposal:

Safely discard potentially contaminated or no longer usable medical devices as healthcare (clinical) waste in accordance with local healthcare guidelines, country and government legislation or policy. Separation, recycling or disposal of packaging material shall follow local country and government legislation on packaging and packaging waste, where applicable. If there is no current legislation, pack them in a perforating waste/sharps disposal container and dispose of them in hospital waste.

Prosthetic Screw

Basic UDI-DI Information:

The following table lists the Basic UDI-DI information for the devices described in this IFU.

Product	Catalog Number	Included Parts	Basic UDI-DI Number
Prosthetic Screw 10PK, Unigrip	PS10PK	PS-001	+D990PS10PK0
Prosthetic Screw 10PK, Straumann	PS10PK-ST	PS-002	+D990PS10PK-ST0
Prosthetic Screw 10PK, TiLobe (Quad Drive)	KDTL-PS10PK	PS-003	+D990KDTL-PS10PK0
Prosthetic Screw 10PK, Paltop (IHex)	KDIH-PS10PK	PS-004	+D990KDIH-PS10PK0
Prosthetic Screw 10PK, BioHorizons	BHHD-PS10PK	PS-005	+D990BHHD-PS10PK0
Premium Starter Kit, SDC	PSK	PS-001	+D990PSK0
Starter Kit, SDC	SK	PS-001	+D990SK0
Recharge Kit w/ Tall Spare Parts, SDC	RK	PS-001	+D990RK0
Recharge Kit w/ POC, SDC	RKPOC	PS-001	+D990RKPOC0
Separable Fastener Assembly 10-pack with Drill kit	SFA10PKDK	PS-001	+D990SFA10PKDK0
Premium Starter Kit, Straumann	PSK-ST	PS-002	+D990PSK-ST0
Starter Kit, Straumann	SK-ST	PS-002	+D990SK-ST0
Recharge Kit w/ Tall Spare Parts, Straumann	RK-ST	PS-002	+D990RK-ST0
Recharge Kit w/ POC, Straumann	RKPOC-ST	PS-002	+D990RKPOC-ST0
Separable Fastener Assembly 10-pack with Drill kit, Straumann	SFA10PKDK-ST	PS-002	+D990SFA10PKDK-ST0
Premium Starter Kit, TiLobe	KDTL-PSK	PS-003	+D990KDTL-PSK0
Starter Kit, TiLobe	KDTL-SK	PS-003	+D990KDTL-SK0
Recharge Kit w/ Tall Spare Parts, TiLobe	KDTL-RK	PS-003	+D990KDTL-RK0
Recharge Kit w/ POC, TiLobe	KDTL-RKPOC	PS-003	+D990KDTL-RKPOC0
Separable Fastener Assembly 10-pack with Drill Kit, TiLobe	KDTL-SFA10PKDK	PS-003	+D990KDTL-SFA10PKDK0
Premium Starter Kit, Paltop	KDIH-PSK	PS-004	+D990KDIH-PSK0
Starter Kit, Paltop	KDIH-SK	PS-004	+D990KDIH-SK0
Recharge Kit w/ Tall Spare Parts, Paltop	KDIH-RK	PS-004	+D990KDIH-RK0
Recharge Kit w/ POC, Paltop	KDIH-RKPOC	PS-004	+D990KDIH-RKPOC0
Separable Fastener Assembly 10-pack with Drill kit, Paltop	KDIH-SFA10PKDK	PS-004	+D990KDIH-SFA10PKDK0
Premium Starter Kit, BioHorizons	BHHD-PSK	PS-005	+D990BHHD-PSK0
Starter Kit, BioHorizons	BHHD-SK	PS-005	+D990BHHD-SK0
Recharge Kit w/ Tall Spare Parts, BioHorizons	BHHD-RK	PS-005	+D990BHHD-RK0
Recharge Kit w/ Tall Spare Parts	BHHD-RKPOC	PS-005	+D990BHHD-RKPOC0
Separable Fastener Assembly 10-pack with Drill kit, BioHorizons	BHHD-SFA10PKDK	PS-005	+D990BHHD-SFA10PKDK0
Premium Starter Kit, TSV	ZVTS-PSK	PS-007	+D990ZVTS-PSK0
Premium Starter Kit, Low Profile	ZVLP-PSK	PS-008	+D990ZVLP-PSK0
Recharge Kit w/ Tall Spare Parts, TSV	ZVTS-RK	PS-007	+D990ZVTS-RK0
Recharge Kit w/ Tall Spare Parts, Low Profile	ZVLP-RK	PS-008	+D990ZVLP-RK0
Recharge Kit w/ POC, TSV	ZVTS-RKPOC	PS-007	+D990ZVTS-RKPOC0
Recharge Kit w/ POC, Low Profile	ZVLP-RKPOC	PS-008	+D990ZVLP-RKPOC0

Prosthetic Screw

Separable Fastener Assembly 10PK - Drill Kit Included, TSV	ZVTS-SFA10PKDK	PS-007	+D990ZVTS-SFA10PKDK0
Separable Fastener Assembly 10PK - Drill Kit Included, Low Profile	ZVLP-SFA10PKDK	PS-008	+D990ZVLP-SFA10PKDK0

Validity:

Upon publication of these instructions for use, all previous versions are superseded.

Availability:

Some items of Smart Denture Conversions are not available in all countries.


























Warranty:

Please visit www.SmartDentureConversions.com for the most up to date warranty information.

Prosthetic Screw

Symbols Glossary:

The following Symbols may be present on the device labeling or in information accompanying the device. Refer to the device labeling or accompanying information for the applicable symbols.

	Manufacturer		Date of Manufacture
	Use By Date		Serial Number
	Batch Code		Catalog Number
	Unique Device Identifier		Medical Device
	CE Mark		UK Representative
	United Kingdom Conformity Assessment Mark		United Kingdom Conformity Assessment Mark with Approved Body Number
	European Representative		Non-Sterile
	Comes Sterilized		Comes Sterilized using Aseptic Processing
	Comes Sterilized using Ethylene Oxide Processing		Comes Sterilized using Irradiation Processing
	Comes Sterilized using Dry Heat Processing		Do Not Resterilize
	Do Not Reuse		Consult Instructions for Use
	Keep Dry		Keep Away from Sunlight
RX Only	For Prescription Use Only		Caution, Consult Accompanying Documents